

the most common candidates to be formed on a hot filament by thermal dissociation.

### Acknowledgments

One of the authors, E. W. Blauth, died shortly before completion of this work. We shall keep him in fond memory. This work was performed under

the terms of the agreement on association between the Max-Planck-Institut für Plasmaphysik and EURATOM.

### References

1. E. W. Blauth and G. Venus, Proc. Int. Vac. Congr. 3rd **2II**, 523 (1965).
2. H. J. Liebl, J. Appl. Phys. **38**, 5277 (1967).

## Automation of a Residual Gas Analyzer on a Time Shared Computer

H. F. Winters, D. L. Raimondi, P. M. Grant, and D. C. Clarke

*IBM Research Laboratory, San Jose, California 95114*  
(Received 28 July 1971)

The automation of a quadrupole mass spectrometer-residual gas analyzer (RGA) on a time-shared IBM 1800 Data Acquisition and Control Computer is described. The RGA, which is used to determine the partial pressures of various gases in a vacuum system, may be operated up to a maximum data rate of  $\sim 20\,000$  points/sec (pps) in an interleaved manner with multiple slow scanning ( $\leq 20$  pps) instruments. We will review in detail the hardware and software considerations regarding the design and subsequent interfacing of the instrument to the computer. Methods for initiating data collection, entering experimental parameters, and analyzing the experimental data, such as spectrum plotting, smoothing, peak location, mass identification, and calculation of the partial pressures by using a least squares approximation to fit mass peaks, will be discussed with examples.